

Equipment	Centrifuge Non-Refrigerated		
Business unit	OR Tambo & Chris Hani		
Laboratory	Canzibe		
Req. #			
Date of Req.			
Specifications	Detail	Compliant	Non Compliant
Drive	Brushless induction		
Control	Microprocessor		
Speed	3000-5000 rpm		
Programmability	Pre set and most recent, visual digital		
Balance detection	Electronic with alarm and cut off		
Dimensions	Bench top: Approx.:300x 400x 600mm		
Run Time	Set and continuous		
Decibels	≤ 60		
Manufacture standards	ISO 13485/9001		
Refrigerated	No		
Number of tubes per adaptor/ per rotor configuration	48 (rotor 4 head adaptor with 12 tubes per bucket)		
Tube size	4-13ml		
Rotor	Swing head		
Adaptor bucket safety	Aerolsol-tight		
Lid	lock with safety release		
Delivery period	2-4 weeks		
Warranty period	Minimum 1 year		
Operators manual	Yes		
Calibration certification	Yes		
Preventative maintenance	Minimum of 1 per year within warranty period with calibration certification		

Laboratory manager	Winisile Bentswana
Signature	PP Bentswana
Date	04/06/2021
QA	N. DUNGA
Signature	ND
Date	04/06/2021