

|                  |                                           |            |
|------------------|-------------------------------------------|------------|
| <b>Equipment</b> | <b>Portable / Free standing Fume Hood</b> |            |
| Business Unit    | NMAL                                      |            |
| Laboratory       | Histology                                 |            |
| Date             |                                           | 06/07/2022 |

|                                                                    |                                           |
|--------------------------------------------------------------------|-------------------------------------------|
| <b>Specification</b>                                               | <b>Detail</b>                             |
| Bench top / mobile                                                 | Yes                                       |
| Dimension                                                          | 1200 mm x 600 mm x 1000 mm                |
| Construction (chemically resistant)                                | Stainless steel                           |
| Front / sash design                                                | folding front sash                        |
| Air velocity                                                       | Compliant                                 |
| Application: volatile, flammable, acids, xylene, noxious chemicals | Compliant                                 |
| Filters                                                            | Carbon activated                          |
| Warranty                                                           | 1 year minimum                            |
| Installation                                                       | Commissioning with compliance certificate |
| Fan control                                                        | Adjustable                                |
| Lighting                                                           | 1000 lux fluorescent lighting             |

|                    |                                                                                     |
|--------------------|-------------------------------------------------------------------------------------|
| <b>Lab manager</b> | <b>Alrika de Beer</b>                                                               |
| Sign               |  |
| Date               | 06/07/2022                                                                          |

|                 |                                                                                      |
|-----------------|--------------------------------------------------------------------------------------|
| <b>NMAL QAC</b> | <b>Xolisa Madikane</b>                                                               |
| Sign            |  |
| Date            | 07/07/2022                                                                           |