

Equipment	Centrifuge Non-Refrigerated		
Business unit	OR Tambo and Chris Hani		
Laboratory	Queenstown		
Req. #			
Dte of Req.			
Specifications	Detail	Vendor Compliance/non compliant	
Drive	Brushless Induction		
Control	Microprocessor		
Speed	3000-5000 rpm		
Programmability	Pre set and most recent, visual digital		
Balance detection	Electronic with alarm and cut off		
Dimensions	Approx.: 300x 400x 600mm		
Run Time	Set and continuous		
Decibels	≤ 60		
Manufacture standards	ISO 13485/9001		
Refrigerated	No		
Number of tubes per adaptor/ per rotor configuration	48 (rotor 4 head adaptor with 12 tubes per bucket)		
Tube size	4-13ml		
Rotor	Swing head		
Adaptor bucket safety	Aerosol-tight		
Lid	lock with safety release		
Delivery period	2-4 weeks		
Warranty period	Minimum 1 year		
Operators manual	Yes		
Calibration certification	Yes		

Laboratory manager	P. Prusent
Signature	<i>PP</i>
Date	
QA	N. Dilinga
Signature	<i>ND</i>
Date	