



Equipment	Waterbath
Business Unit	Alfred No 8 Joe Goobi
Laboratory	Taylor Bequest Mt Fletcher
Req. #	
Specification	Detail
Footprint	Bench Top
Construction	Stainless Steel
Temperature adjustment	Variable
Operating temperature requirements	37°C
Temperature tolerance range	Less than $\pm 2.0^{\circ}\text{C}$
Digital temperature display	Yes
Temperature out of range alarm	Optional
Circulatory pump	Yes
Timer	No
Maximum dimensions (Bench space) (L x W x H in mm)	L = 450mm W = 450mm H = 200mm
Tank Capacity (in litres)	Maximum 5L
Weight when filled (in kg)	Maximum 30kg
Waterbath cover (Lid)	Yes
Test tube rack	Minimum 4 x 5mL tubes
Drain or ease of cleaning	Yes
Safety	Low water level sensor
Warranty period	Minimum 1 year
Delivery period	6 weeks
Aftersales service	No
Operator's manual	English
Electronic voltage	SA Compliant
Lab manager	A. Krump
Signature	
Date	21.06.2023
QA	N. Godfrey
Signature	
Date	21 June 2023